



Certificate

Herewith we, OMRON HEALTHCARE Co., Ltd. certify HEM-9200T is substantially equivalent to HEM-7130 (Japanese model) and M6 AC(HEM-7322-E), both clinically validated. It employs the same algorithm and the same components as the predicate device for determining blood pressure. HEM-7130 (Japanese model) and M6 AC (HEM-7322-E) fulfill requirements of International Protocol revision 2010 by European Society of Hypertension.*1

[Reference]

1. Dabl Educational Trust. Blood pressure monitors-validation, papers, and reviews. Available at: <http://www.dableducational.org>. [Accessed 29 March 2016].

Very Truly Yours,

April 1, 2016

Manabu Hamamoto

General Manager

Bio-Information System Development Department1

Product Development Strategy HQ